

Plan of the intercomparison of distributed calibration objects in relative humidity	Issue 2	
Approved by Håkan Källgren	Date 2026-06-25	Page 1 (5)

Content

Content	1
ILC humidity 2026:1	2
Intercomparison of relative humidity sensors.....	2
Relative humidity ranges	2
Proficiency testing provider (PT).....	2
Participants welcome to this intercomparison.....	2
Description of the values included in the intercomparison.	2
Description of probe 1 Vaisala	3
Description of probe 2 Testo	3
Detailed documented instructions	3
Time schedule and quality check.....	3
Statistical analyses that will be used	4
Reporting	5
Damaged PT/ILC item	5
Price for participation	5

Plan of the intercomparison of distributed calibration objects in relative humidity	Issue 2	
Approved by Håkan Källgren	Date 2026-06-25	Page 2 (5)

ILC humidity 2026:1

Intercomparison of relative humidity sensors

The instruments will be transported between participating laboratories.

Relative humidity ranges

10-90 %rh at temperature 23 °C,

You may participate even if you cannot calibrate all levels.

The calibrations shall be done in air.

The above concept of intercomparison was decided by an advisory group related to the calibration of humidity.

No subcontractors are involved in the intercomparison.

Proficiency testing provider (PT)

Swedish Metrology and Quality AB (SMQ) is organizing this intercomparison on calibration of the following objects.

Participants welcome to this intercomparison.

Participation is open for three categories of laboratories:

- Accredited laboratories
- Laboratories that will apply for accreditation.
- Laboratories that want to evaluate their Calibration and Measurement Capability (CMC).

Due to practical reasons the number of participants is limited to minimum 8 and maximum 20.

Description of the values included in the intercomparison.

The intercomparison will start and end with both sensors calibrated at the reference laboratory which also will define the reference values. Its CMC levels for uncertainty of the reference laboratory are given in the tables below as well as the possibly reference uncertainty when calibrating the object.

Reference laboratory will be JV Norway.

Plan of the intercomparison of distributed calibration objects in relative humidity	Issue 2	
Approved by Håkan Källgren	Date 2026-06-25	Page 3 (5)

Description of probe 1 Vaisala

Range	Calibration point	Reference laboratory CMC values *	Possible expanded reference uncertainty **
°C	%rh	%rh	%rh
23	10	0,4	0,6
23	30	0,5	0,8
23	50	0,5	0,8
23	70	0,6	0,9
23	90	0,8	1,2
*CMC values from the KCDB database at bimp.org			
**The values are expressed as expanded uncertainty, U with 95 % confidence level and include possible drift during circulation of the objects among the participating laboratories.			

Description of probe 2 Testo

Range	Calibration point	Reference laboratory CMC values *	Possible expanded reference uncertainty **
°C	%rh	%rh	%rh
23	10	0,4	0,6
23	30	0,5	0,8
23	50	0,5	0,8
23	70	0,6	0,9
23	90	0,8	1,2
*CMC values from the KCDB database at bimp.org			
**The values are expressed as expanded uncertainty, U with 95 % confidence level and include possible drift during circulation of the objects among the participating laboratories.			

Detailed documented instructions

Detailed technical instructions will be sent to the participants who have registered to the ILC together with the reporting form as an excel document.

Time schedule and quality check

After registration of interested laboratories a plan for circulating of the calibration objects and a time schedule will be worked out and sent to all of them.

The calibration objects will eventually be transported by different means in participating countries.

Plan of the intercomparison of distributed calibration objects in relative humidity	Issue 2	
Approved by Håkan Källgren	Date 2026-06-25	Page 4 (5)

Immediately at receiving and sending the objects each participant shall inform the organizer (SMQ) by e-mail about the status of the objects. If there are any signs of impact (for example marks or scratches) a photograph shall be sent to the organizer to decide how to proceed and to inform the next participant.

Each participant will have access to the object for maximum 10 working days, including transport, and should use its own method for calibration.

Laboratories freely decide which uncertainty they want to state in the protocol, accredited laboratories can declare their CMC-values, or a value estimated for the actual measurement condition.

After finishing calibration, the objects shall be sent to the following participant on the transportation list using the same parcel they arrived in.

Original data from the calibration shall be sent to the organizer immediately after finishing the calibrations. This is preferably done by e-mailing the filled excel-reporting file. By a fast delivery, the organizer gets control that everything is as expected. It also helps to detect eventual arising problems in time.

Statistical analyses that will be used

The organiser will calculate reference values that will be used in the calculations as described in ISO/IEC 17043:2010 annex B3 presenting En-values (formula B5)

Plan of the intercomparison of distributed calibration objects in relative humidity	Issue 2	
Approved by Håkan Källgren	Date 2026-06-25	Page 5 (5)

Reporting

Participants shall send their final calibration certificate to the organiser within one week after the calibrations are finished in form of a pdf-file in a mail message.

The excel-reporting document that will be sent to all laboratories shall indicate important environment conditions.

At the end of the intercomparison a draft report will be returned to the participants within 2 weeks from the time when the last participant has reported its results in a calibration certificate.

The participants are encouraged to comment on the draft report within one week after receiving it.

If a participant does not follow the described reporting rules without giving reasonable explanations the organizer tries to extract the relevant content for the comparison. If this is not possible the results will be excluded from the report.

A participant may decide to withdraw from the exercise. This might be caused by problems detected during or after having performed the measurements. However, the withdrawal in this case must be announced to the organizer before the draft report is distributed to all participants.

The participant may appeal to the full report if there should be major faults in the report.

The intercomparison report will list up all participating laboratories. However, each will be treated anonymously, and its result will only be identified by a code that is send to each participant in a separate e-mail.

Damaged PT/ILC item

The participant shall immediately inform the organizer in case of any anomalies or other detected problem to allow him to take appropriate actions.

Price for participation

Total price includes the basic price and the additional cost of calibration for the probes.

Amount of calibration technicians in laboratory	Basic price	Total price for calibrating 1 probe	Total price for calibrating 2 probes
3 or less	750 EUR	1330 EUR	1910 EUR
More than 3	980 EUR	1560 EUR	2140 EUR

The laboratories shall pay the transport costs to next laboratory.

If the laboratory decides not to fulfil its part of the agreement after they have registered, they shall still pay the basic price.